



## CENTRE FOR NANOSCIENCE & NANOTECHNOLOGY

(South Campus, Block-II, Sec-25)  
PANJAB UNIVERSITY, CHANDIGARH-160014 (INDIA)

NSNT/3338-3342

Tel. No: + 91-172-2541741

Date: 04/05/2017

FAX No: +91-172-2783336

**Speed post/Registered post**

Dear Sir,

Please quote **Technical and Financial bids (in two separate envelopes)** for the supply of following item(s) to the Principal Investigator, Centre for Nanoscience and Nanotechnology, Panjab University, Chandigarh as per specifications: -

Sr. No.	Generic Name – Spin Coating Unit (Spin Coater) Quantity required: 1 No
1.	<b><u>Specifications</u></b> <ul style="list-style-type: none"><li>i. Make: MBraun/SUSS Microtec/Novocontrol/Ossila/Laurell</li><li>ii. Spin coating unit should be suitable for use in glove box. Preferably should have provisions for nitrogen purging.</li><li>iii. Should be suitable for both circular or rectangular substrates held by vacuum suction/automatic suction on a rotational platform.</li><li>iv. RPM: 8000 rpm or higher for uniform film coating.</li><li>v. Rotational platform for variable substrates of different dimensions as per requirements of the end user. Additional stages (chucks) should be supplied according to the substrate dimensions if a particular rotational platform (chuck) is not suitable for substrates with other sizes. Preferably should be able to hold both circular and rectangular substrates.</li><li>vi. The unit should be all inclusive supplied with all necessary accessories. Suitable vacuum pump should also be supplied along with the unit, if it is not inbuilt.</li><li>vii. Electrical requirement: Up to 240 V/ 50 Hz or as per Indian standard, single phase.</li><li>viii. Warranty: Minimum two years.</li><li>ix. Installation within 21 days after arrival of the instrument.</li></ul>

The quotations must reach in sealed envelopes by **19.05.2017 at 5.00 p.m.** along with your other terms and conditions of supply, if any.

**LAST DATE OF RECEIPT OF QUOTATIONS: 19.05.2017 AT 5.00 P.M.**

**\*OPENING OF QUOTATIONS: 22.05.2017 AT 3.30 P.M.**

**(\*If it is a holiday, next working day will be considered)**

**Note: -** The quotation must reach by hand/registered post/speed post on or before **19.05.2017 by 5.00 p.m.** at the following address.

**The Principal Investigator,  
Centre for Nanoscience and Nanotechnology, Panjab University,  
Science Block- II, Sector-25, University South Campus, Chandigarh-160014.**

**- MOST IMPORTANT-**

1. The bidders are requested to attach the EMD as demand draft of 2 % of the total value of the quotation/proforma invoice in the name of 'The coordinator, Centre for Nanoscience and nanotechnology, Panjab University Chandigarh', **without EMD financial bids will not be entertained**. However, EMD is not required for quotations of total value of Rs. 1.00 Lakh or below.
2. Please quote technical and financial bids separately (in two separate envelopes) with EMD.
3. Panjab University, does not take responsibility for any postal delay in delivery of registered/speed post or lost in transit of quotation.
4. We have been exempted from paying central excise duty to vide govt. of India notification no. 10/97-central excise dated March 1, 1997 and is valid up to 31-8-2020.
5. The rate should be quoted both in words and figures in **financial bid only**.
6. Conditional and unsigned quotation will not be accepted.
7. The supply of consignment be commenced/made within 30 days of the issue of supply order.
8. All quoted rates should be for Panjab University and the firm should quote the rate of all taxes.
9. No payment will be made on proforma invoice.
10. The quotations shall not contain corrections, erasers, and overwriting.
11. The undersigned reserves the right to accept or reject **any quotation without assigning any reason. Lack of workmanship in the product may warrant rejection of the product even after delivery.**
12. **Special discount for the educational institution, University teaching department may be mentioned.**
13. Quotation shall be opened on **22-05-2017 at 3.30 p.m.** and **you may depute your representative at the time of opening of quotation.**
14. The quotation (technical and financial bid with EMD) in a sealed envelope giving our/your reference no/date of quotation should be sent after affixing the required postage stamps. The quotation may be sent by speed post /registered post (as far as possible) / by hand.
15. The present rate of S.T. applicable on the articles should be clearly mentioned.
16. Earnest money/security deposit/any other sums of the tenderers lying with the University in connection with and other tender/case will not be considered against this tender.
17. Technical specifications are intended to be descriptive only and not restrictive. The bidder may substitute alternatives with identical standards, brand names and/or catalogue numbers in its bids, provided that it demonstrates to the purchaser's satisfaction.

18. The bidders shall provide a list of reputed institutions/universities/research laboratories in India where their equipment had been installed.

**Mention clearly whether you meet the Technical Specifications (yes/no)**

<b>Sr. No.</b>	<b>Generic Name of Items</b>	<b>Specify Yes/No</b>
<b>1.</b>	<b>Spin Coating Unit (Spin Coater)</b>	
	<p style="text-align: center;"><b><u>Specifications</u></b></p> <ul style="list-style-type: none"><li>i. Make: MBraun/SUSS Microtec/Novocontrol/Ossila/Laurell</li><li>ii. Spin coating unit should be suitable for use in glove box. Preferably should have provisions for nitrogen purging.</li><li>iii. Should be suitable for both circular or rectangular substrates held by vacuum suction/automatic suction on a rotational platform.</li><li>iv. RPM: 8000 rpm or higher for uniform film coating.</li><li>v. Rotational platform for variable substrates of different dimensions as per requirements of the end user. Additional stages (chucks) should be supplied according to the substrate dimensions if a particular rotational platform (chuck) is not suitable for substrates with other sizes. Preferably should be able to hold both circular and rectangular substrates.</li><li>vi. The unit should be all inclusive supplied with all necessary accessories. Suitable vacuum pump should also be supplied along with the unit, if it is not inbuilt.</li><li>vii. Electrical requirement: Up to 240 V/ 50 Hz or as per Indian standard, single phase.</li><li>viii. Warranty: Minimum two years.</li><li>ix. Installation within 21 days after arrival of the instrument.</li></ul>	

Yours faithfully,



(Dr. Jadab Sharma)

Principal Investigator,

SERB-ECR Grant,

Centre for Nanoscience & Nanotechnology, Panjab University